

File : Q:\GCMS-08 Minimal\DATA\20220518\14- -JP.D
Operator : Organic QC Lab
Acquired : 18 May 22 07:51 pm using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: C-137S-TP
Misc Info : 22'344'5-HexaPCB; in iso
Vial Number: 59



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

