

File :D:\MassHunter\GCMS\1\data\20211206\17- ALR.D
Operator : Organic QC Lab
Acquired : 06 Dec 2021 15:40 using AcqMethod Hy_PCB_10cm_S2.M
Instrument : GCMS 8
Sample Name: C-143S-TP
Misc Info : 2,2',3,4,5,6'-Hexachlorobiphenyl in Isooctane
Vial Number: 83



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

