

File : L:\QC Data\GCMS-09 Minimal\DATA\08252022\06 - SAK.D
Operator :
Acquired : 25 Aug 2022 13:15 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split Time 12 min .M
Instrument : GCMS 9
Sample Name : C-146S-TP
Misc Info : 2,2,3,4,5,5 Hexachlorobiphenyl in Iso
Vial Number : 7



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

