

File :Q:\GCMS-08 Minimal\DATA\20211130\22- = JP ].D  
Operator : Organic QC Lab  
Acquired : 30 Nov 2021 17:28 using AcqMethod Hy\_Pest\_24cm\_S1.M  
Instrument : GCMS 8  
Sample Name: C-157S-TP  
Misc Info : 233'44'5-Hexa-PCB; in iso  
Vial Number: 59



Column: DB-5, 30m, 0.25ID, 0.25um  
Oven: 80c, 2m, 30c/m to 320, final 2min  
GC para: Pulse split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj, GCMS, Inj 280c

