

File :Q:\GCMS-08 Minimal\DATA\20210527\30- .D
Operator : Organic QC Lab
Acquired : 27 May 2021 18:54 using AcqMethod Hy_PAH_24cm_S1.M
Instrument : GCMS 8
Sample Name: C-166S (221051665)
Misc Info : 2,3,4,4',5,6-Hexa-PCB; in isoctane
Vial Number: 51



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

