

File :Q:\GCMS-08 Minimal\DATA\20220810\25 =JP=.D  
Operator : Organic QC Lab  
Acquired : 10 Aug 22 08:40 pm using AcqMethod Hy\_Pest\_15cm\_S5.M  
Instrument : GCMS 8  
Sample Name: C-167S  
Misc Info : 23'44'55'-Hexa-PCB; in iso  
Vial Number: 5



Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

