

File :Q:\GCMS-06 Minimal\DATA\120519\02-.D
Operator :
Acquired : 05 Dec 2019 15:16 using AcqMethod CHICK_2019_2S_2uL.M
Instrument : GCMS 6
Sample Name: C-221S-TP
Misc Info : Aroclor 1221 in Isooctane
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

