

File :D:\MassHunter\GCMS\1\DATA\03252201\02- ALR.D
Operator : Organic QC department
Acquired : 25 Mar 2022 10:30 using AcqMethod KERLA-2Split-2uL.M
Instrument : 5977B MSD
Sample Name: C-221S
Misc Info : Aroclor 1221 in Isooctane
Vial Number: 27



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

