

File : L:\QC Data\GCMS-06 Minimal\DATA\082423\17 - sak.D
Operator : Organic QC Lab
Acquired : 24 Aug 2023 21:56 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: C-232S-H
Misc Info :
Vial Number: 108

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

