

File :Q:\GCMS-06 Minimal\DATA\120721\44--DC.D
Operator : Organic QC Lab
Acquired : 08 Dec 21 08:07 am using AcqMethod CHICK_2019_2S_2uL.M
Instrument : GCMS 6
Sample Name: C-232S
Misc Info : Aroclor 1232 in Isooctane
Vial Number: 58



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 2:1, 2uL inj.; GC/MS; INJ 270c

