

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-07 Minimal\DATA\04052301
... \09- KM.D
Operator : Organic QC department
Instrument : 5977B MSD
Acquired : 05 Apr 2023 09:13 using AcqMethod KERLA-2017.M
Sample Name: C-242S-H
Misc Info : Aroclor 1242 in Hexane



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

