

File :Q:\GCMS-08 Minimal\DATA\20190515\BSB\04- .D  
Operator : JP/KM  
Acquired : 15 May 2019 17:57 using AcqMethod KERLA-2017-100uL draw speed.M  
Instrument : GCMS 8  
Sample Name: C-254S  
Misc Info : Aroclor 1254 in Isooctane  
Vial Number: 52



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

