

File :Q:\GCMS-06 Minimal\DATA\112719\BSB\07-.D  
Operator : KM  
Acquired : 27 Nov 2019 12:52 using AcqMethod CHICK\_2017\_S50.M  
Instrument : GCMS 6  
Sample Name: C-254S-M-28.5X  
Misc Info : Aroclor 1254 in MeOH  
Vial Number: 48



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

