

File :M:\GCMS-06 Minimal\DATA\122821\25-ALR.D  
Operator : Organic QC Lab  
Acquired : 28 Dec 2021 15:12 using AcqMethod CHICK\_2019\_2S\_2uL.M  
Instrument : GCMS 6  
Sample Name: C-260S-H-10X  
Misc Info : Aroclor 1260 in Hexane  
Vial Number: 105



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

