

File :Q:\GCMS-08 Minimal\DATA\20210920\32- JP=.D
 Operator : Organic QC Lab
 Acquired : 20 Sep 2021 21:44 using AcqMethod Hy_VOC-SV_15cm_S50.M
 Instrument : GCMS 8
 Sample Name: C-CCSEC
 Misc Info : PCBs in acetone
 Vial Number: 56



Column: DB-5, 30m, 0.25ID, 0.25um
 Oven: 80c, 2m, 30c/m to 320, final 2min
 GC para: Pulse split 40psi 0.5m, 10psi constant flow
 Split 20:1, 1ul inj, GCMS, Inj 280c

