

File :D:\MassHunter\GCMS\1\data\07021901\03- .D  
 Operator : KD  
 Acquired : 02 Jul 2019 10:37 using AcqMethod KERLA-2Split-2uL.M  
 Instrument : 5977B MSD  
 Sample Name: C-CS-02  
 Misc Info : in Isooctane  
 Vial Number: 54

Column: DB-5, 30m, 0.25 ID, 0.25um  
 Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
 GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
 Split 20:1, 1ul inj.; GC/MS; INJ 250c.

