

File :Q:\GCMS-04 Minimal\DATA\022122\35- YX.D  
Operator : Organic QC Lab  
Acquired : 22 Feb 2022 4:58 using AcqMethod VOC-100SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name: CLP-003-R  
Misc Info : Purgeable Organic Matrix  
Vial Number: 16



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

