

File :Q:\GCMS-08\DATA\20190122\07- D
Operator : JJP / KM
Acquired : 22 Jan 2019 17:39 using AcqMethod KERLA-2017-Splitless_200EM.M
Instrument : GCMS 8
Sample Name: CLP-024R-160X
Misc Info : CLP Pesticide Standard in Isooctane
Vial Number: 7



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

