

File :D:\MassHunter\GCMS\1\DATA\11182001\24- KD.D
Operator : Organic QC department
Acquired : 18 Nov 20 03:23 pm using AcqMethod KERLA-2017_S50.M
Instrument : 5977B MSD
Sample Name: CLP-BNS
Misc Info in DCM
Vial Number: 41

Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

