

File :Q:\GCMS-08 Minimal\DATA\20211028\12- KM.D  
Operator : Organic QC Lab  
Acquired : 28 Oct 2021 16:16 using AcqMethod Hy\_VOC-SV\_15cm\_S50.M  
Instrument : GCMS 8  
Sample Name: CLP-HC-A 8)  
Misc Info : Acid Composite Mixture in MeCl2  
Vial Number: 6



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

