

File :Z:\QC Data\GCMS-07 Minimal\DATA\01102001\30-.D
Operator : Organic QC department
Acquired : 10 Jan 2020 19:47 using AcqMethod KERLA-2017_S50.M
Instrument : 5977B MSD
Sample Name: CLP-HC-X1-R1
Misc Info in DCM
Vial Number: 132



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

