

File :K:\QC Data\GCMS-04 Minimal\DATA\041321C\21- RM.D  
 Operator :  
 Acquired : 14 Apr 2021 10:28 using AcqMethod VOC-100SPLIT DECKER.M  
 Instrument : GCMS4  
 Sample Name: CLP-PIPS  
 Misc Info :  
 Vial Number: 22

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
 Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
 GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

