

File : L:\QC Data\GCMS-04 Minimal\DATA\072920\39- ZA.D
Operator : ORGANIC QC DEPT.
Acquired : 30 Jul 2020 3:12 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name : CLP-PS-3
Misc Info :
Vial Number : 62



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

