

File : Q:\GCMS-04 Minimal\DATA\063022\37- RM.D
Operator :
Acquired : 1 Jul 2022 1:52 using AcqMethod VOC-100SPLIT#.M
Instrument : GCMS4
Sample Name: CLP-PS
Misc Info :
Vial Number: 72

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

