



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : LCMS-1                     Location  : P2-E2
Injection Date  : 12/12/2019 6:38:32 AM      Inj       :    2
                                           Inj Volume: 1.000 µl
  
```

```

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/18/2019 4:18:14 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 12/11/2019 1:10:22 PM by SYSTEM
                 (modified after loading)
  
```

```

Method Info    : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
                Reference.
  
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

