



=====
 Acq. Operator : SYSTEM Seq. Line : 10
 Sample Operator : SYSTEM
 Acq. Instrument : LCMS-2 Location : P2-C-10
 Injection Date : 1/23/2025 9:39:38 AM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\012325\012325 2025-01-23 06-49-01
 \Purity_195-315.M
 Last changed : 2/6/2023 2:02:18 PM by SYSTEM
 Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
 Last changed : 2/6/2023 2:02:18 PM by SYSTEM
 Method Info : OPERATOR: Organic QC lab
 General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
 Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
 plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
 , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

