

File :D:\MassHunter\GCMS\1\DATA\060923\17 -SAK.D  
Operator : Organic QC Lab  
Acquired : 09 Jun 2023 09:28 using AcqMethod CHICK\_2019\_VOC-SV\_5min.M  
Instrument : GCMS 6  
Sample Name: D-4629-91-LB-10X  
Misc Info : 10.0 ug/mL  
Vial Number: 20



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

