

File :Q:\GCMS-07 Minimal\DATA\10251901\26-.D
Operator : KD & DC
Acquired : 25 Oct 2019 23:06 using AcqMethod KERLA-2Split-2uL_VOC-SV_W-200EM.M
Instrument : 5977B MSD
Sample Name: D-4929-94
Misc Info : Chlorine Standard in Isooctane
Vial Number: 142



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

