

File : Q:\GCMS-08 Minimal\DATA\20230511\47- [JP].D
Operator : Organic QC Lab
Acquired : 11 May 2023 23:31 using AcqMethod Hy_HC08-HC40_S100.M
Instrument : GCMS 8
Sample Name: D-5442-CR
Misc Info : in cyclohexane
Vial Number: 41



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

