

=====  
Acq. Operator : SYSTEM Seq. Line : 11  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P2-D-10  
Injection Date : 8/10/2020 10:33:10 AM Inj : 1  
Inj Volume : 1.000 µl  
Method : C:\Users\Public\Documents\ChemStation\1\Data\08102001\08032001 2020-08-10  
07-23-13\Purity\_195-315.M (Sequence Method)  
Last changed : 5/8/2020 10:12:19 AM by SYSTEM  
Method Info : OPERATOR: Organic QC lab  
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax  
Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity  
II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

