

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-07 Minimal\DATA\02292301
... \41- KM.D
Operator : Organic QC department
Instrument : 5977B MSD
Acquired : 02 Mar 2023 03:41 using AcqMethod KERLA-VOC-SV_S100.M
Sample Name: D-6379-SRS-R1
Misc Info : D-6379 System Resolution Standard in Heptane



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

