


File :M:\GCMS-04 Minimal\DATA\060421\45- ALR.D  
 Operator :  
 Acquired : 5 Jun 2021 7:44 using AcqMethod VOC-100SPLIT HOT\_28 MIN.M  
 Instrument : GCMS4  
 Sample Name: S-77855-03  
 Misc Info : D7845 Calibration/IS Level 3 in Toluene  
 Vial Number: 33



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
 Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
 GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C



Not ID'd by GC/MS: 1-Butanol (71-36-3), trans-b-Methylstyrene (873-66-5), and Styrene Glycol (93-56-1)