

```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : LCMS-1                     Location  : P1-C5
Injection Date  : 6/15/2021 3:34:22 PM       Inj       :    1
                                                Inj Volume: 1.000 µl
Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl
```



```
Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 6/16/2021 10:33:12 AM by SYSTEM
                (modified after loading)
```

```
Method Info    : Operator: Organic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV
                Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

