

File :Q:\GCMS-08 Minimal\DATA\20221219\38- 2 [JP].D
Operator : Organic QC Lab
Acquired : 19 Dec 22 07:47 pm using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: DIN38414-20-IS
Misc Info : Deca-PCB; in hexane
Vial Number: 33



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

