

File :D:\MassHunter\GCMS\1\data\20220124\02- KM.D
Operator : Organic QC Lab
Acquired : 24 Jan 22 11:16 am using AcqMethod Hy_Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: DRFH-007-SS
Misc Info : 1-Chlorooctadecane in Hexane
Vial Number: 29



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

