

File :M:\GCMS-07 Minimal\DATA\12102101\12-ALR.D
 Operator : Organic QC department
 Acquired : 10 Dec 2021 21:17 using AcqMethod KERLA-2Split-2uL_VOC-SV_W-200EM.M
 Instrument : 5977B MSD
 Sample Name: DRH-TX-003-FCS
 Misc Info : Aromatic Fractionation Check in Pentane
 Vial Number: 100



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
 Split 20:1, 1µl inj.; GC/MS; Injector 270°C

