

File : Q:\GCMS-08 Minimal\DATA\20190925\39- .D
Operator : ZA / KM / JP
Acquired : 26 Sep 2019 05:14 using AcqMethod KERLA-2017_S100.M
Instrument : GCMS 8
Sample Name: DRO-AK-102-LCS-10X
Misc Info : #2 Diesel Fuel in Acetone
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

