

File : Q:\GCMS-08 Minimal\DATA\20230324\15- (JP).D
Operator : Organic QC Lab
Acquired : 24 Mar 23 05:44 pm using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: DRO-AK-102-SS
Misc Info : o-Terphenyl; in acet
Vial Number: 6



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

