

File :Z:\QC Data\GCMS-07 Minimal\DATA\12051901\02-.D  
Operator : Organic QC department  
Acquired : 05 Dec 2019 11:55 using AcqMethod KERLA-2017\_S50.M  
Instrument : 5977B MSD  
Sample Name: DRO-AK-102AA  
Misc Info : in DCM  
Vial Number: 11



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

