

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    4
Acq. Instrument : LCMS-1                      Location  : P1-D9
Injection Date  : 8/16/2022 7:24:29 AM       Inj       :    1
                                           Inj Volume: 1.000 µl

```

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

```

Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 11/5/2021 2:37:17 PM by SYSTEM
Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.

```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

