



=====  
Acq. Operator : SYSTEM Seq. Line : 17  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P1-C-05  
Injection Date : 8/23/2021 12:21:33 PM Inj : 1  
Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\08232101\08232101 2021-08-23 07-15-52  
\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%  
Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system  
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

