



Acq. Operator : SYSTEM
 Sample Operator : SYSTEM
 Acq. Instrument : LCMS-2
 Injection Date : 7/23/2020 3:47:30 PM
 Seq. Line : 3
 Location : P1-E-10
 Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\07232001\07202001 2020-07-23 15-08-38
 \Purity_195-315.M
 Last changed : 5/8/2020 10:12:19 AM by SYSTEM
 Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
 Last changed : 5/8/2020 10:12:19 AM by SYSTEM
 Method Info : OPERATOR: Organic QC lab
 General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
 Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
 plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system
 , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

