



=====

Acq. Operator	: SYSTEM	Seq. Line	: 19
Sample Operator	: SYSTEM		
Acq. Instrument	: LCMS-2	Location	: P1-B-07
Injection Date	: 1/17/2022 3:15:53 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\01172201\01132201 2022-01-17  
09-34-52\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax  
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity  
II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Basic Red 9 @ 100ug/mL in MeOH

Additional Info : Peak(s) manually integrated

### Current Chromatogram(s)

