



=====

Acq. Operator : SYSTEM	Seq. Line :
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P1-C6
Injection Date : 3/15/2023 12:53:15 PM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 3.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\031523\031523101 2023-03-15 10-19-24  
\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Disperse Yellow 1 in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

