

Sample Name : DYE-057S

Acq. Operator : SYSTEM

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Injection Date : 3/3/2025 8:56:44 AM

Seq. Line :

Location : P2-A-03

Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\030325\022825 2025-03-03 07-59-09\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : FD & C Red 3 MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

