

=====

Acq. Operator : SYSTEM	Seq. Line : 34
Acq. Instrument : LCMS-1	Location : P2-A9
Injection Date : 11/15/2022 4:42:29 PM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 5.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 11/5/2021 2:37:17 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 11/9/2022 8:35:37 AM by SYSTEM  
(modified after loading)

Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Sample Info : Direct Red 28 in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

