

Sample Name : DYE-112S-R1

**AccuStandard®**
Leader in Analytical Reference Standards

Acq. Operator : SYSTEM

Seq. Line :

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P2-E-08

Injection Date : 3/11/2025 2:56:20 PM

Inj : 1

Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 5.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\031125\031025 2025-03-11 07-41-49
\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Ponceau SX in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

