

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    5
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P2-C-05
Injection Date  : 3/19/2020 12:15:16 PM      Inj       :    1
                                           Inj Volume: 1.000 µl

```

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\03192001\02192001 2020-03-19 10-58-54 \Purity\_195-315.M

Last changed : 12/3/2019 1:25:24 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 3/17/2020 3:54:55 PM by SYSTEM  
(modified after loading)

Method Info : General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

