



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :   14
Acq. Instrument : LCMS-1                               Location  :   P1-E8
Injection Date  : 2/28/2022 12:18:23 PM              Inj       :    1
                                                    Inj Volume: 1.000 µl
    
```

```

Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 11/5/2021 2:37:17 PM by SYSTEM
Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple
                  MS Detector. No UV Reference.
    
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

