

File :Q:\GCMS-06\DATA\030819\23-.D  
Operator : D.C  
Acquired : 09 Mar 2019 03:12 using AcqMethod CHICK\_2017\_S200.M  
Instrument : GCMS 6  
Sample Name: EN-12916-01  
Misc Info : EN12916 Cal Std in Heptane  
Vial Number: 140



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 200:1, 1uL inj.; GC/MS; INJ 270c

